

**MANUFACTURER:**

Ranyák György  
H-1094 Budapest, Tompa u. 15/b. - Hungary

**DESCRIPTION OF THE DEVICE:**

Nasal secretion suction device

**TYPE:**

*Suction method:* with vacuum cleaner,  
*Raw material:* glass or plastic  
*Components:* suction devices, connecting tube, connecting bell  
*Accessories:* soft head, dachshund shape

**INTENDED USE:**

Apparatus for removing the nasal secretions of a child unable to blow his nose, the vacuum source of which is a household vacuum cleaner

**MODEL:**

**ANGEL-VAC**

**CLASSIFICATION:**

Class IIa, Rule 2 and 11 according to Annex IX of the MDD

**DECLARATION:**

We herewith declare that the above mentioned products meet the provisions of the Council Directive 93/42/EEC for medical devices and other standards. All supporting documentation is retained under the premises of the manufacturer.

**NOTIFIED BODY:**

CE-CERTISO Ltd.  
H-2092 Budakeszi, Erdő utca 101.  
NB ID number: 2409

**EC CERTIFICATE:**

No: 145061-21-01-07  
Requirement: 93/42/EEC Annex V  
Expires: 25 May 2024

Budapest, 07-01-2021

  
György Ranyák  
manager

